

Calibration Certificate *SL***Accelerometer System**TYPE ~~8002K~~

Accelerometer Model 8076K

SN C106761

Amplifier/Coupler Model 5020

SN C34162

Range $\pm$ 250 g*Sensitivity at 100 Hz, 10 g*rms 10.00 mV/g*
at Hz, g*rms

Mounted Resonant Frequency 37.3 kHz

159.15

Transverse Sensitivity 2.0 %

*.010498**10.01 - 0.001 g**
*9.999 mV/g**Mounting Torque 24 ± 2 lbf-in (2.7 ± 0.2 Nm)

All measurements at ambient conditions of 21°C (70°F) and 50% R.H.

*g = 9.807 m/s²✓ Frequency Response *159.15*

Frequency (Hz)	20	50	100	200	500	1k	2k	5k	8k	10k
Deviation (%)	-0.9	-0.1	+0.0	-0.1	-0.4	-0.1	+0.1	+0.4	-0.3	-0.7

This accelerometer was calibrated using a back-to-back comparison technique against a Kistler Working Standard. Kistler Working Standards are periodically calibrated against a Kistler Reference Standard System

±0.06 which in turn is periodically recertified by the National Institute of Standards and Technology. The calibration of all Kistler acceptance test instrumentation is in conformance with MIL-STD-45662A.

NIST Traceability

Working Standard

Reference Standard

Accelerometer

Model 8002K

SN C28623

Model 808K1

SN 1263

Charge Amplifier

Model 5020

SN C31242

Model 561T

SN 251

NIST Test Report Number 822/250337

AUG 15 1996

By Mark Thomas

Date 08-15-96